

Plasma Medicine

Alexander Fridman | Gary Friedman

An abstract, glowing visualization of plasma, showing intricate, swirling patterns of light in shades of purple, blue, and white against a dark background. The light appears to be contained within a dark, curved vessel or chamber.

 WILEY

Plasma Medicine

Plasma Medicine

Alexander Fridman and Gary Friedman
Drexel University, Philadelphia, PA, USA



A John Wiley & Sons, Ltd., Publication

This edition first published 2013
© 2013 John Wiley & Sons, Ltd

Registered office

John Wiley & Sons Ltd, The Atrium, Southern Gate, Chichester, West Sussex, PO19 8SQ, United Kingdom

For details of our global editorial offices, for customer services and for information about how to apply for permission to reuse the copyright material in this book please see our website at www.wiley.com.

The right of the author to be identified as the author of this work has been asserted in accordance with the Copyright, Designs and Patents Act 1988.

All rights reserved. No part of this publication may be reproduced, stored in a retrieval system, or transmitted, in any form or by any means, electronic, mechanical, photocopying, recording or otherwise, except as permitted by the UK Copyright, Designs and Patents Act 1988, without the prior permission of the publisher.

Wiley also publishes its books in a variety of electronic formats. Some content that appears in print may not be available in electronic books.

Designations used by companies to distinguish their products are often claimed as trademarks. All brand names and product names used in this book are trade names, service marks, trademarks or registered trademarks of their respective owners. The publisher is not associated with any product or vendor mentioned in this book. This publication is designed to provide accurate and authoritative information in regard to the subject matter covered. It is sold on the understanding that the publisher is not engaged in rendering professional services. If professional advice or other expert assistance is required, the services of a competent professional should be sought.

The publisher and the author make no representations or warranties with respect to the accuracy or completeness of the contents of this work and specifically disclaim all warranties, including without limitation any implied warranties of fitness for a particular purpose. This work is sold with the understanding that the publisher is not engaged in rendering professional services. The advice and strategies contained herein may not be suitable for every situation. In view of ongoing research, equipment modifications, changes in governmental regulations, and the constant flow of information relating to the use of experimental reagents, equipment, and devices, the reader is urged to review and evaluate the information provided in the package insert or instructions for each chemical, piece of equipment, reagent, or device for, among other things, any changes in the instructions or indication of usage and for added warnings and precautions. The fact that an organization or Website is referred to in this work as a citation and/or a potential source of further information does not mean that the author or the publisher endorses the information the organization or Website may provide or recommendations it may make. Further, readers should be aware that Internet Websites listed in this work may have changed or disappeared between when this work was written and when it is read. No warranty may be created or extended by any promotional statements for this work. Neither the publisher nor the author shall be liable for any damages arising herefrom.

Library of Congress Cataloging-in-Publication Data

Fridman, Alexander A., 1953–

Plasma medicine / Alexander Fridman, Gary Friedman.

p. ; cm.

Includes index.

ISBN 978-0-470-68970-7 (cloth) – ISBN 978-0-470-68969-1 (paper)

I. Friedman, Gary (Gary G.) II. Title.

[DNLM: 1. Biomedical Engineering. 2. Plasma Gases–therapeutic use. 3. Plasma Gases–pharmacology. 4. Wound Healing. QT 36] 610.28–dc23

2012025754

A catalogue record for this book is available from the British Library.

Print ISBN: Cloth: 9780470689707 Paper: 9780470689691

Typeset in 10/12pt Times by Aptara Inc., New Delhi, India

Contents

Preface	xv
Acknowledgements	xvii
1 Introduction to Fundamental and Applied Aspects of Plasma Medicine	1
1.1 Plasma medicine as a novel branch of medical technology	1
1.2 Why plasma can be a useful tool in medicine	4
1.3 Natural and man-made, completely and weakly ionized plasmas	5
1.4 Plasma as a non-equilibrium multi-temperature system	7
1.5 Gas discharges as plasma sources for biology and medicine	9
1.6 Plasma chemistry as the fundamental basis of plasma medicine	13
1.7 Non-thermal plasma interaction with cells and living tissues	14
1.8 Applied plasma medicine	15
2 Fundamentals of Plasma Physics and Plasma Chemistry for Biological and Medical Applications	19
2.1 Elementary plasma generation processes	19
2.1.1 Classification of ionization processes	19
2.1.2 Direct and stepwise ionization by electron impact	21
2.1.3 Other ionization mechanisms in plasma-medical discharges	23
2.1.4 Mechanisms of electron-ion recombination in plasma	24
2.1.5 Elementary processes of negative ions in plasma-medical systems: Electron attachment and detachment	26
2.1.6 Heterogeneous ionization processes, electron emission mechanisms	29
2.2 Excited species in plasma medicine: Excitation, relaxation and dissociation of neutral particles in plasma	31
2.2.1 Vibrational excitation of molecules by electron impact	31
2.2.2 Rotational and electronic excitation of molecules by electron impact	33
2.2.3 Plasma generation of atoms and radicals: Dissociation of molecules by electron impact	34
2.2.4 Distribution of non-thermal plasma discharge energy between different channels of excitation of atoms and molecules	35
2.2.5 Relaxation of plasma-excited species: Vibrational-translational (VT) relaxation	36
2.2.6 Vibrational energy transfer between molecules: VV-relaxation	37
2.2.7 Relaxation of rotational and electronic excitation in plasma-medical systems	38
2.3 Elementary plasma-chemical reactions of excited neutrals and ions	39
2.3.1 Rate coefficient of reactions of excited molecules in plasma	39
2.3.2 Efficiency α of excitation energy in overcoming activation energy of chemical reactions of plasma-generated active species	40

2.3.3	Fridman-Macheret α -model for chemical reactions of plasma-generated active species	40
2.3.4	Ion-molecular polarization collisions and reactions in plasma	43
2.3.5	Ion-molecular chemical reactions of positive and negative ions	45
2.4	Plasma statistics, thermodynamics, and transfer processes	46
2.4.1	Statistical distributions in plasma: Boltzmann distribution function, Saha equation, and dissociation equilibrium	46
2.4.2	Complete and local thermodynamic equilibrium in plasma	48
2.4.3	Thermodynamic functions of quasi-equilibrium thermal plasma systems	49
2.4.4	Non-equilibrium statistics of thermal and non-thermal plasmas	50
2.4.5	Non-equilibrium statistics of vibrationally excited molecules in plasma and Treanor distribution	51
2.4.6	Transfer processes in plasma: Electron and ion drift	52
2.4.7	Transfer processes in plasma: Diffusion of electrons and ions	54
2.4.8	Transfer processes in plasma: Thermal conductivity	55
2.4.9	Radiation energy transfer in plasma	56
2.5	Plasma kinetics: Energy distribution functions of electrons and excited atoms and molecules	58
2.5.1	Electron energy distribution functions (EEDF) in plasma: Fokker-Planck kinetic equation	58
2.5.2	Relation between electron temperature and the reduced electric field	60
2.5.3	Non-equilibrium vibrational distribution functions of plasma-excited molecules, Fokker-Planck kinetic equation	61
2.5.4	Vibrational distribution functions of excited molecules in plasma controlled by VV-exchange and VT-relaxation processes	64
2.5.5	Kinetics of population of electronically excited states in plasma	65
2.5.6	Macrokinetics of chemical reactions of vibrationally excited molecules	66
2.6	Plasma electrodynamics	68
2.6.1	Ideal and non-ideal plasmas	68
2.6.2	Plasma polarization, Debye shielding of electric field in plasma	69
2.6.3	Plasmas and sheaths	69
2.6.4	Electrostatic plasma oscillations, Langmuir plasma frequency	72
2.6.5	Penetration of slow-changing fields into plasma: Skin-effect in plasma	72
2.6.6	Plasma magneto-hydrodynamics: Generalized Ohm's law, Alfvén velocity and magnetic Reynolds' number	73
2.6.7	High-frequency conductivity and dielectric permittivity of plasma	75
2.6.8	Propagation of electromagnetic waves in plasma	77
2.6.9	Plasma absorption and reflection of electromagnetic waves: Bouguer law and critical electron density	78
3	Selected Concepts in Biology and Medicine for Physical Scientists	81
3.1	Molecular basis of life: Organic molecules primer	81
3.1.1	Essential primer on bonds and organic molecules	81
3.1.2	Main classes of organic molecules in living systems	84
3.2	Function and classification of living forms	96
3.2.1	What is life: Functionality of living forms	96
3.2.2	Major classification of living forms	96

3.3	Cells: Organization and functions	97
3.3.1	Primary cell components	98
3.3.2	Transport across cell membranes	105
3.3.3	Cell cycle and cell division	109
3.3.4	Cellular metabolism	110
3.3.5	Reactive species in cells and living organisms	115
3.4	Overview of anatomy and physiology	124
3.4.1	Tissues	124
3.4.2	The body covering: The integumentary system	130
3.4.3	Circulatory system	132
3.4.4	Immune system	139
3.4.5	Digestive system	144
3.4.6	The nervous system	149
3.4.7	The endocrine system	155
3.4.8	The muscular and skeletal systems	157
3.4.9	The respiratory system	160
3.4.10	The excretory system	163
4	Major Plasma Discharges and their Applicability for Plasma Medicine	165
4.1	Electric breakdown and steady-state regimes of non-equilibrium plasma discharges	165
4.1.1	Townsend mechanism of electric breakdown, Paschen curves	165
4.1.2	Streamer or spark breakdown mechanism	168
4.1.3	Meek criterion of streamer formation and streamer propagation models	171
4.1.4	Streamers and microdischarges	172
4.1.5	Interaction of streamers and microdischarges	173
4.1.6	Steady-state regimes of non-equilibrium electric discharges	174
4.1.7	Discharge regime controlled by electron-ion recombination	175
4.1.8	Discharge regime controlled by electron attachment	175
4.1.9	Non-thermal discharge regime controlled by diffusion of charged particles to the walls	176
4.2	Glow discharge and its application to biology and medicine	177
4.2.1	Glow discharge structure	177
4.2.2	Current-voltage characteristics of DC discharges	180
4.2.3	Townsend dark discharge	180
4.2.4	Current-voltage characteristics of the cathode layer	182
4.2.5	Abnormal, subnormal and obstructed regimes of glow discharges	184
4.2.6	Positive column of glow discharge	186
4.2.7	Atmospheric pressure glow discharges and applications in plasma medicine	186
4.2.8	Resistive barrier discharge (RBD) as modification of APG discharges	188
4.2.9	Atmospheric pressure micro glow-discharges	188
4.2.10	Hollow-cathode glow discharge and hollow-cathode APG microplasma	191
4.3	Arc discharge and its medical applications	191
4.3.1	Major types of arc discharges	191
4.3.2	Cathode and anode layers of arc discharges and spots	193
4.3.3	Positive column of high-pressure arcs	196
4.3.4	Steenbeck – Raizer ‘channel’ model of positive column of arc discharges	198

4.3.5	Configurations of arc discharges and their applicability to plasma medicine	199
4.3.6	Gliding arc discharge as a powerful source of non-equilibrium plasma	204
4.4	Radio-frequency and microwave discharges in plasma medicine	208
4.4.1	Generation of thermal plasma in radio-frequency discharges	208
4.4.2	Atmospheric-pressure microwave discharges and their biomedical applications	210
4.4.3	Non-thermal RF discharges: CCP and ICP coupling	213
4.4.4	Non-thermal RF-CCP discharges at moderate pressure regime	215
4.4.5	Low-pressure CCP RF discharges	217
4.4.6	Low-pressure RF magnetron discharges for surface treatment	219
4.4.7	Low-pressure non-thermal ICP RF discharges in cylindrical coil	219
4.4.8	Planar-coil and other configurations of low-pressure non-thermal RF-ICP discharges	222
4.4.9	Non-thermal RF atmospheric pressure plasma jets as surface-treatment device	224
4.4.10	Atmospheric-pressure non-thermal RF plasma microdischarges: Plasma needle	226
4.4.11	Non-thermal low-pressure microwave and other wave-heated discharges	227
4.4.12	Non-equilibrium microwave discharges of moderate and elevated pressures: Energy-efficient plasma source of chemically active species	229
4.5	Coronas, DBDs, plasma jets, sparks and other non-thermal atmospheric-pressure streamer discharges	232
4.5.1	Corona and pulsed corona discharges	232
4.5.2	Dielectric-barrier discharges (DBDs)	234
4.5.3	Special modifications of DBD: Surface, asymmetric, packed bed and ferroelectric discharges	235
4.5.4	OAUGDP as quasi-homogeneous DBD (APG) modification	237
4.5.5	Electronically stabilized DBD in APG discharge mode	238
4.5.6	Arrays of DBD-based microdischarges and kilohertz-frequency microdischarges	239
4.5.7	Floating-electrode dielectric barrier discharge (FE-DBD)	241
4.5.8	Micro- and nanosecond pulsed uniform FE-DBD plasma	243
4.5.9	Spark discharges	243
4.5.10	Pin-to-hole spark discharge (PHD), thermal microplasma source-generating ROS and NO for medical applications	247
4.5.11	Atmospheric-pressure cold helium microplasma jets and plasma bullets	250
4.5.12	Propagation of plasma bullets in long dielectric tubes and splitting and mixing of plasma bullets	252
4.6	Discharges in liquids	253
4.6.1	General features of electrical discharges in liquids in relation to their biomedical applications	253
4.6.2	Mechanisms and characteristics of plasma discharges in water	254
4.6.3	Physical kinetics of water breakdown: Thermal breakdown mechanism	254
4.6.4	Non-thermal short pulse electrostatic (electrostriction) water breakdown	256
4.6.5	Nanosecond pulse breakdown and plasma generation in liquid water without bubbles	261
4.6.6	Nanosecond-pulse uniform cold plasma in liquid water without bubbles: Analysis and perspectives for biomedical applications	263

5	Mechanisms of Plasma Interactions with Cells	269
5.1	Main interaction stages and key players	269
5.2	Role of plasma electrons and ions	272
5.2.1	Selection of biological targets and plasma generation methods	272
5.2.2	Comparison of direct DBD plasma treatment to indirect treatment with and without ion flux	272
5.2.3	Effect of gas composition on antibacterial efficacy of direct DBD	274
5.2.4	Effect of positive and negative ions in nitrogen corona discharge	276
5.3	Role of UV, hydrogen peroxide, ozone and water	276
5.3.1	Effect of UV in DBD treatment	276
5.3.2	Effect of hydrogen peroxide	277
5.3.3	Effect of ozone	279
5.3.4	Effects of water and its amount	280
5.4	Biological mechanisms of plasma interaction for mammalian cells	281
5.4.1	Intracellular ROS as key mediators of plasma interaction with mammalian cells	281
5.4.2	DNA damage and repair as a consequence of DBD plasma treatment	283
5.4.3	Effect of the cell medium in plasma interaction with mammalian cells	286
5.4.4	Crossing the cell membrane	289
6	Plasma Sterilization of Different Surfaces and Living Tissues	293
6.1	Non-thermal plasma surface sterilization at low pressures	293
6.1.1	Direct application of low-pressure plasma for biological sterilization	293
6.1.2	Effect of low-pressure plasma afterglow on bacteria deactivation	294
6.2	Surface microorganism inactivation by non-equilibrium high-pressure plasma	295
6.2.1	Features of atmospheric-pressure air plasma sterilization	295
6.2.2	Kinetics of atmospheric-pressure plasma sterilization	295
6.2.3	Cold plasma inactivation of spores: <i>Bacillus cereus</i> and <i>Bacillus anthracis</i> (anthrax)	297
6.2.4	Atmospheric-pressure air DBD plasma inactivation of <i>Bacillus cereus</i> and <i>Bacillus anthracis</i> spores	298
6.2.5	Decontamination of surfaces from extremophile organisms using non-thermal atmospheric-pressure plasma	301
6.2.6	Plasma sterilization of contaminated surgical instruments: Prion proteins	303
6.3	Plasma species and factors active for sterilization	304
6.3.1	Direct effect of charged particles in plasma sterilization	304
6.3.2	Biochemical effect of plasma-generated electrons in plasma sterilization	304
6.3.3	Bio-chemical effect of plasma-generated negative and positive ions	306
6.3.4	Sterilization effect of ion bombardment	307
6.3.5	Sterilization effect of electric fields related to charged plasma particles	307
6.3.6	Effect of plasma-generated active neutrals: ROS and RNS	308
6.3.7	Effect of plasma-generated active neutrals: OH-radicals and ozone	309
6.3.8	Effects of plasma-generated active neutrals: Hydrogen peroxide (H ₂ O ₂)	311
6.3.9	Contribution of plasma-generated heat and temperature to plasma sterilization	311
6.3.10	Effect of UV radiation	312

6.4	Physical and biochemical effects of atmospheric-pressure air plasma on microorganisms	313
6.4.1	Direct and indirect effects of non-thermal plasma on bacteria	313
6.4.2	FE-DBD experiments demonstrating higher effectiveness of direct plasma treatment	315
6.4.3	Surface versus penetrative plasma sterilization	317
6.4.4	Apoptosis versus necrosis	320
6.5	Animal and human living tissue sterilization	320
6.5.1	Direct FE-DBD for living tissue treatment	320
6.5.2	Direct FE-DBD plasma source for living tissue sterilization	321
6.5.3	Toxicity (non-damaging) analysis of direct plasma treatment of living tissue	322
6.6	Generated active species and plasma sterilization of living tissues	324
6.6.1	Physico-chemical in vitro tissue model: Production and delivery in tissue of active species generated in plasma	324
6.6.2	FE-DBD plasma system for analysis of deep tissue penetration of plasma-generated active species	325
6.6.3	Deep tissue penetration of plasma-generated active species	327
6.7	Deactivation/destruction of microorganisms due to plasma sterilization: Are they dead or just scared to death?	329
6.7.1	Biological responses of <i>Bacillus stratosphericus</i> to FE-DBD plasma treatment	329
6.7.2	FE-DBD plasma treatment of <i>Bacillus stratosphericus</i>	331
6.7.3	Analysis of deactivation/destruction of <i>Bacillus stratosphericus</i> due to non-thermal plasma sterilization	336
6.7.4	Bottom line for plasma physicists: Plasma sterilization can lead to VBNC state of microorganisms	338
7	Plasma Decontamination of Water and Air Streams	339
7.1	Non-thermal plasma sterilization of air streams	339
7.1.1	Direct sterilization versus application of filters	339
7.1.2	Pathogen detection and remediation facility	340
7.1.3	The dielectric barrier grating discharge (DBGD) applied in the PDRF	343
7.1.4	Rapid and direct plasma deactivation of airborne bacteria in the PDRF	343
7.1.5	Phenomenological kinetic model of non-thermal plasma sterilization of air streams	345
7.1.6	Kinetics and mechanisms of rapid plasma deactivation of airborne bacteria at the PDRF	346
7.2	Direct and indirect effects in non-thermal plasma deactivation of airborne bacteria	347
7.2.1	Major sterilization factors	347
7.2.2	PDRF: experimental procedure	348
7.2.3	PDRF: experimental results	351
7.3	Non-thermal plasma in air-decontamination: Air cleaning from SO ₂ and NO _x	353
7.3.1	Plasma cleaning of industrial SO ₂ emissions	353
7.3.2	SO ₂ oxidation to SO ₃ using relativistic electron beams	354
7.3.3	SO ₂ oxidation to SO ₃ using continuous and pulsed corona discharges	354
7.3.4	Plasma-stimulated liquid-phase chain oxidation of SO ₂ in droplets	355
7.3.5	Plasma-catalytic chain oxidation of SO ₂ in clusters	358

7.3.6	Simplified mechanism and energy balance of the plasma-catalytic chain oxidation of SO ₂ in clusters	359
7.3.7	Plasma-stimulated combined oxidation of NO _x and SO ₂ in air; simultaneous industrial exhaust gas cleaning from nitrogen and sulfur oxides	360
7.4	Non-thermal plasma decontamination of air from volatile organic compound (VOC) emissions	361
7.4.1	General features of non-thermal plasma treatment of VOC emissions in air	361
7.4.2	Mechanisms and energy balance of treatment of VOC exhaust gases from paper mills and wood processing plants	362
7.4.3	Removal of acetone and methanol from air using pulsed corona discharge	364
7.4.4	Removal of dimethyl sulfide from air using pulsed corona discharge	364
7.4.5	Removal of α -pinene from air using pulsed corona discharge	368
7.4.6	Treatment of paper mill exhaust gases using wet pulsed corona discharge	369
7.4.7	Non-thermal plasma decontamination of diluted large-volume emissions of chlorine-containing VOC	370
7.4.8	Non-thermal plasma removal of elemental mercury from coal-fired power plants and other industrial air-based off-gases	376
7.4.9	Mechanism of non-thermal plasma removal of elemental mercury from air streams	377
7.5	Plasma disinfection and sterilization of water	378
7.5.1	Plasma water disinfection using UV-radiation, ozone and pulsed electric fields	378
7.5.2	Applications of pulsed plasma discharges for water treatment	379
7.5.3	Energy-effective water treatment using pulsed spark discharges	380
7.5.4	Characterization of the pulsed spark discharge system applied for energy-effective water sterilization	381
7.5.5	Analysis of D-value and role of UV radiation in inactivation of microorganisms in water	385
8	Plasma Treatment of Blood	389
8.1	Plasma-assisted blood coagulation	389
8.1.1	General features of plasma-assisted blood coagulation	389
8.1.2	Experiments with non-thermal atmospheric-pressure plasma-assisted in vitro blood coagulation	389
8.1.3	In-vivo blood coagulation using FE-DBD plasma	391
8.1.4	Mechanisms of non-thermal plasma-assisted blood coagulation	391
8.1.5	Influence of protein activity	393
8.2	Effect of non-thermal plasma on improvement of rheological properties of blood	395
8.2.1	Control of low-density-lipoprotein (LDL) cholesterol and blood viscosity	395
8.2.2	Plasma-medical system for DBD plasma control of blood properties	395
8.2.3	DBD plasma control of whole blood viscosity (WBV)	397
8.2.4	DBD plasma effect on improvement of rheological properties of blood	400
9	Plasma-assisted Healing and Treatment of Diseases	403
9.1	Wound healing and plasma treatment of wounds	403
9.1.1	Wounds and healing processes	403
9.1.2	Treatment of wounds using thermal and nitric-oxide-producing plasmas	407
9.1.3	Experience with other thermal discharges	416

9.2	Treatment of inflammatory dysfunctions	418
9.2.1	Examples of anti-inflammatory treatment by Plazon	418
9.2.2	Pin-to-hole microdischarge for ulcerative colitis treatment	419
9.3	Plasma treatment of cancer	422
9.3.1	Observations of cultured malignant cells	424
9.3.2	Non-thermal plasma treatment of explanted tumors in animal models	426
9.4	Plasma applications in dentistry	428
9.4.1	Brief overview of structure of teeth and dental health	429
9.4.2	Recent promising results of plasma applications in dentistry	431
9.5	Plasma surgery	433
10	Plasma Pharmacology	435
10.1	Non-thermal plasma treatment of water	435
10.2	Deionized water treatment with DBD in different gases: Experimental setup	436
10.2.1	Changing the working gas and sample degassing	436
10.3	Deionized water treatment with DBD in different gases: Results and discussion	438
10.4	Enhanced antimicrobial effect due to organic components dissolved in water	442
10.4.1	Setup and sample preparation	442
10.4.2	Comparison of DBD treated water, PBS and NAC solutions	444
10.5	Summary	446
11	Plasma-assisted Tissue Engineering and Plasma Processing of Polymers	447
11.1	Regulation of biological properties of medical polymer materials	447
11.1.1	Tissue engineering and plasma control of biological properties of medical polymers	447
11.1.2	Wettability or hydrophilicity of medical polymer surfaces for biocompatibility	448
11.2	Plasma-assisted cell attachment and proliferation on polymer scaffolds	448
11.2.1	Attachment and proliferation of bone cells on polymer scaffolds	448
11.2.2	DBD plasma effect on attachment and proliferation of osteoblasts cultured over PCL scaffolds	449
11.3	Plasma-assisted tissue engineering in control of stem cells and tissue regeneration	451
11.3.1	About plasma-assisted tissue regeneration	451
11.3.2	Control of stem cell behavior on non-thermal plasma-modified polymer surfaces	452
11.3.3	Plasma-assisted bioactive liquid micro-xerography	453
11.4	Plasma-chemical polymerization of hydrocarbons and formation of thin polymer films	453
11.4.1	Biological and medical applications of plasma polymerization	453
11.4.2	Mechanisms and kinetics of plasma polymerization	454
11.4.3	Initiation of polymerization by dissociation of hydrocarbons in plasma volume	455
11.4.4	Heterogeneous mechanisms of plasma-chemical polymerization of C ₁ /C ₂ hydrocarbons	456
11.4.5	Plasma-initiated chain polymerization: Mechanisms of plasma polymerization of MMA	457
11.4.6	Plasma-initiated graft polymerization	458
11.4.7	Formation of polymer macro-particles in non-thermal plasma in hydrocarbons	459
11.4.8	Specific properties of plasma-polymerized films	460
11.4.9	Electric properties of plasma-polymerized films	460

11.5	Interaction of non-thermal plasma with polymer surfaces	461
11.5.1	Plasma treatment of polymer surfaces	461
11.5.2	Major initial chemical products created on polymer surfaces during non-thermal plasma interaction	463
11.5.3	Formation kinetics of main chemical products in pulsed RF treatment of PE	464
11.5.4	Kinetics of PE treatment in continuous RF discharge	466
11.5.5	Non-thermal plasma etching of polymer materials	466
11.5.6	Contribution of electrons and UV radiation in plasma treatment of polymer materials	467
11.5.7	Interaction of chemically active heavy particles generated in non-thermal plasma with polymer materials	469
11.5.8	Synergetic effect of plasma-generated active particles and UV radiation with polymers	471
11.5.9	Aging effect in plasma-treated polymers	471
11.5.10	Plasma modification of wettability of polymer surfaces	472
11.5.11	Plasma enhancement of adhesion of polymer surfaces	474
11.5.12	Plasma modification of polymer fibers and polymer membranes	475
	References	483
	Index	497

Preface

Plasma medicine has inspired the last decade of the authors' professional activities at Drexel Plasma Institute of the Drexel University. Plasma medicine is a very exciting and new multidisciplinary branch of modern science and technology. Even the term 'plasma medicine' has only been in existence since the start of the 21st century. Plasma medicine embraces physics required to develop novel plasma discharges relevant for medical applications, medicine to apply the technology not only in vitro but also in vivo testing and, last but not least, biology to understand the complicated biochemical processes involved in plasma interaction with living tissues.

While an understanding of the mechanisms by which non-thermal plasma interacts with living systems has begun to emerge only recently, a significant number of journal publications and even reviews focused on plasma medicine have appeared during the last 5–10 years. Several prestigious journals have published special issues dedicated to the topic, the new *Plasma Medicine* journal has been recently inaugurated, multiple world symposiums have created special sessions in this new field and plasma medical workshops have been organized in the USA, Germany, France, Korea, Japan, China and other countries. Four successful International Conferences on Plasma Medicine (ICPM) took place during the last 7 years in the USA, Germany and France; the 5th ICPM is planned to be held in Japan. Finally, the International Society on Plasma Medicine has been organized to coordinate the efforts of physicist, chemists, biologists, engineers, medical doctors and representatives of the industry working in this new field.

Despite the tremendous interest in plasma medicine, no single monograph has published in this field. There is no book where recent developments in plasma medicine, both technological and scientific, are described in a fashion accessible to the highly interdisciplinary audience of doctors, physicists, biologists, chemists and other scientists, university students and professors, engineers and medical practitioners. This is exactly the goal of the present *Plasma Medicine* book. The book is written for numerous scientists and medical practitioners, students, professors and industrial professionals who are involved today in plasma medicine.

When writing the book, we kept in mind the multidisciplinary nature of the field of plasma medicine. Physicists, chemists and engineers should be able to learn the different terminology of their biologist and medical practitioner partners, and vice versa. The book is beneficial to sides and should promote more effective development of the field of plasma medicine. The subject of plasma medicine has recently been included in the academic curriculum of universities, and we hope that the book will be helpful in this regard to students (as well as professors) involved in plasma-medical education.

Plasma Medicine consists of 11 chapters; Chapters 1–5 are focused on the **fundamentals of plasma medicine** and Chapter 6–11 are focused on **applied plasma medicine**.

Chapter 1 introduces the subject of plasma medicine. Chapter 2 describes the fundamental physical and chemical processes in plasmas relevant to its interaction with living tissues, providing a basic introduction to plasma medicine. Chapter 3 describes fundamental biology relevant to an understanding of the major principles of plasma interaction with living tissues. This topic covers the basic biological and medical introduction to plasma medicine, and will help physicists and engineers understand that even simple living organisms are much more complicated than electric devices. Chapter 4 describes plasma physics and engineering of the systems and devices relevant for medical applications. This chapter covers physical, chemical and engineering aspects of major electric discharges used for plasma–medical applications. In chapter 5, a description of the

biophysical and biochemical mechanisms of plasma interaction with living tissues is provided. This chapter enables an understanding of the kinetics of plasma interaction with eukaryotic and prokaryotic cells, starting from gas phase and surface processes stimulated by active plasma species and including the consequent biochemical processes inside the cells.

Chapters 6 and 7 describe plasma sterilization of inanimate surfaces, as well as plasma sterilization of water and air. These chapters cover multiple applications of different low-pressure and atmospheric-pressure non-thermal discharges for disinfection and sterilization of different surfaces (e.g. medical instruments, food, space-crafts etc.); natural, drinking and industrial water; and large-volume air flows. Chapter 8 is focused on plasma-induced cauterization and blood coagulation. Plasma control of blood composition and relevant plasma treatment of blood diseases is also discussed in this chapter. Chapter 9 describes plasma treatment and healing of different wounds and diseases, in particular, plasma abatement of skin, dental and internal infections, treatment and healing of wounds and plasma treatment of oncological (cancerous), gastrointestinal, cardiovascular and other diseases. Chapter 10 describes plasma pharmacology, which suggests preliminary plasma treatment of water or special organic or inorganic solutions. These plasma-treated solutions can then be utilized for sterilization or healing purposes. The last chapter is focused on basic aspects of plasma-medical tissue engineering. This topic covers major modern aspects of plasma treatment of biomaterials and plasma-supported tissue engineering. This very important topic of applied plasma medicine is not directly related to plasma interaction with living tissue. This part of modern plasma medicine is very interesting, significant and relatively better developed (in particular, by our colleagues from University of Bari) than other branches; only a concise review of the subject is provided in this book.

Instructors can access PowerPoint files of the illustrations presented within this text, for teaching, at <http://booksupport.wiley.com>.

Alexander Fridman and Gary Friedman
Philadelphia

Acknowledgements

The authors gratefully acknowledge the support of their families, support of Drexel Plasma Institute (DPI) by John and Chris Nyheim and Kaplans family, support from leaders of Drexel University (President John Fry, Provost Mark Greenberg, Dean Joe Hughes, and senior vice-provost Deborah Crawford). The authors greatly appreciate support of their plasma-medical research by the National Institute of Health, National Science Foundation, US Department of Defense (especially DARPA and TATRC), W.M. Keck Foundation, Coulter Foundation, NASA and USDA; as well as the support of our industrial sponsors, especially Johnson & Johnson and GoJo.

For stimulating discussions on the topic of plasma medicine and assistance in development of the book, the authors gratefully acknowledge their colleagues and friends from Drexel University, especially Professors Danil Dobrynin, Gregory Fridman, Young Cho, Ari Brooks, Jane Azizkhan-Clifford, Richard Hamilton, Donna Murasko and Banu Onoral as well as Dr Terry Freeman from Jefferson University, Dr Steve Davis from University of Miami, Dr Alexander Gutsol from Chevron, Professor Ken Blank from Temple University, Professor Richard Satava from the University of Washington (formerly with DARPA) and many wonderful graduate students. Special thanks are addressed to Kirill Gutsol for assistance with illustrations.

1

Introduction to Fundamental and Applied Aspects of Plasma Medicine

1.1 Plasma medicine as a novel branch of medical technology

New ideas bring new hopes: plasma medicine is definitely one of those. Recent developments in physics and engineering have resulted in many important medical advances. The various medical technologies that have been widely described in the existing literature include applications of ionizing radiation, lasers, ultrasound, magnetism, and others. Plasma technology is a relative newcomer to the field of medicine. Very recent exponential developments in physical electronics and pulsed power engineering have promoted consequent significant developments in non-thermal atmospheric-pressure plasma science and engineering. Space-uniform and well-controlled cold atmospheric-pressure plasma sources have become a reality, creating the opportunity to safely and controllably apply plasma to animal and human bodies. This has instigated the creation of a novel and exciting area of medical technology: plasma medicine.

Experimental work conducted at several major universities, research centers, and hospitals around the world over the last decade demonstrates that non-thermal plasma can provide breakthrough solutions to challenging medical problems. It is effective in sterilization of different surfaces including living tissues, disinfects large-scale air and water streams, deactivates dangerous pathogens including those in food and drinks, and is able to stop serious bleeding without damaging healthy tissue. Non-thermal plasma can be directly used to promote wound healing and to treat multiple diseases including skin, gastrointestinal, cardiovascular, and dental diseases, as well as different forms of cancer. It has also proven effective in the treatment of blood, controlling its properties. Non-thermal discharges have also proven to be very useful in the treatment of different biomaterials and in tissue engineering, tissue analysis and diagnostics of diseases. Research indicates that non-thermal plasma may prove to be useful in pharmacology by changing properties of existing drugs and creating new medicines. Non-thermal plasma, developed recently due to the rapid progress in electronics, is clearly a promising new tool which should be provided to medical doctors to resolve medical problems. Plasma medicine, the subject of this book, is a source of great interest today.

When talking about the novel plasma sources which it is possible to apply to human and animal bodies, as well as for the treatment of cells and tissues in detailed biomedical experiments, we have to stress the *safety* and *controllability* of these novel plasma devices. As an example, the floating-electrode dielectric barrier discharge (FE-DBD) plasma source widely used for medical applications, in particular in Drexel University,



Figure 1.1 Non-thermal short-pulsed 40 kV FE-DBD plasma sustained directly between a dielectric-coated electrode and a human body (see color plate).

applies c. 30–40 kV directly to the human body (see one of the authors of this book in Figure 1.1). Obviously, safety is the main issue in this case. Of no less importance is the controllability of the plasma parameters. The uniform cold atmospheric-pressure plasmas as well as some other plasma-medical devices developed recently can be effectively controlled; this is important not only for prescribing specific doses of medical treatment, but also for investigation of the mechanisms of plasma-medical treatment. Without detailed understanding of physical, chemical, and biomedical mechanisms, plasma tools have little chance of successful application in medicine.

Non-thermal plasma is very far from thermodynamic equilibrium, which is discussed below. Such strongly non-equilibrium medium can be very ‘creative’ in its interactions with biomolecules. As first demonstrated in the 1950s by Stanley Miller (see Figure 1.2) and his colleagues from the University of Chicago, plasma is even able to generate amino acids from methane and inorganics. It is very much possible that plasma, being a strongly non-equilibrium and multi-parametric medium, can even be responsible for the creation of life itself. Recent experiments prove that controllable changes of DNA after non-thermal plasma treatment are very sensitive to plasma parameters. This explains the great importance of the controllability of plasma parameters and a deep understanding of mechanisms for successful progress of plasma-medical science. The success of plasma medicine requires a detailed understanding of physical, chemical, and biomedical mechanisms of the strongly non-equilibrium plasma interaction with cells and living tissues. Without a fundamental understanding, plasma medicine is at risk of become a modernized medieval magic (see Figure 1.3).

Plasma medicine is a multidisciplinary branch of modern science and technology. It embraces physics (required to develop novel plasma discharges relevant for medical applications), medicine (to apply the technology for not only in vitro but also in vivo testing), and last but not least biology (to understand the complicated biochemical processes involved in plasma interaction with living tissues). While an understanding of the mechanisms by which non-thermal plasma can interact with living systems has begun to emerge only recently, a significant number of original journal publications and even reviews have appeared since the mid-2000s. Several prestigious journals have published special issues dedicated to the plasma medicine, the new *Plasma Medicine* journal has been recently launched, multiple world symposiums have created special sessions in this new field, and plasma-medical workshops have been organized in the USA, Germany, France,



Figure 1.2 In the 1950s, Stanley Miller of the University of Chicago synthesized amino acids in plasma from methane and inorganic compounds (see color plate).



Figure 1.3 International Society for Plasma Medicine (ISPM) signifies crucial importance of deep and detailed research focused on fundamental understanding of physical, chemical and biological bases of plasma medicine.

Korea, Slovakia, and other countries. The most important world forum of plasma-medical research is the International Conferences on Plasma Medicine (ICPM). Four of these biannual conferences have already been successfully organized: ICPM-1 in Corpus Christi, Texas, USA; ICPM-2 in San-Antonia, Texas, USA; ICPM-3 in Greifswald, Germany; and finally ICPM-4 in Orleans, France in 2012. Finally, the International Society on Plasma Medicine was launched this year (2012) to coordinate the efforts of physicist, chemists, biologists, engineers, medical doctors and representatives of the industry in the new field of plasma medicine.

Hopefully, this book will be helpful to this entire and very multidisciplinary group of researchers and industry representatives. Plasma scientists and medical doctors speak different languages; they even have two different meanings for the word 'plasma' itself. Plasma scientists, medical doctors and biologists often have very different approaches to fundamental knowledge as well as practical applicability, but this book recognizes that they are united by a mutual interest in this new field of plasma medicine and by the common idea that development of plasma medicine brings new opportunities for treating human conditions.

1.2 Why plasma can be a useful tool in medicine

While the term 'medicine' in the title of the book does not require a special introduction, the term 'plasma' may require some elucidation (especially for medical practitioners). Plasma is an ionized gas and a distinct fourth state of matter. 'Ionized' means that at least one electron is not bound to an atom or molecule, converting them into positively charged ions. As temperature increases, atoms and molecules become more energetic and the state of matter transforms in the sequence: solid to liquid, liquid to gas and finally gas to plasma, which justifies the label of 'fourth state of matter'.

The free electric charges, electrons and ions make plasma electrically conductive (with magnitudes of conductivity sometimes exceeding that of gold and copper), internally interactive, and strongly responsive to electromagnetic fields. Ionized gas is defined as plasma when it is electrically neutral (electron density is balanced by that of positive ions) and contains a significant number of electrically charged particles, sufficient to affect its electrical properties and behavior. In addition to being important in many aspects of our daily lives, plasmas are estimated to constitute more than 99% of the known universe.

The term 'plasma' was first introduced by Irving Langmuir in 1928 when the multi-component, strongly interacting ionized gas reminded him of blood plasma; the term 'plasma' itself therefore has a strong relation to medicine. This can however be confusing: for example, read the discussions regarding plasma treatment of blood plasma in Chapter 8 of this book. Defining the term plasma, Irving Langmuir wrote: "Except near the electrodes, where there are sheaths containing very few electrons, the ionized gas contains ions and electrons in about equal numbers so that the resultant space charge is very small. We shall use the name **plasma** to describe this region containing balanced charges of ions and electrons". Plasmas occur naturally, but can also be effectively produced in laboratory settings and in industrial or hospital operations, providing opportunities for numerous applications including thermonuclear synthesis, electronics, lasers, fluorescent lamps, cauterization and tissue ablation during surgeries, and many others. We remind the reader that most computer and cell-phone hardware is based on plasma technology, not to forget about the plasma TV. In this book, we will focus on the fundamental and practical aspects of plasma applications to medicine, biology, and related disciplines, which represent today probably the most novel and exciting component of plasma science and engineering. Plasma is widely used in practice today. Generally, plasma offers three major features which are attractive for major practical applications.

1. *Temperatures* and energy densities of some plasma components can significantly exceed those in conventional technologies. These temperatures can easily exceed the level of c. 10 000 K. For example, if melted ceramics are needed to make relevant coatings, requiring temperatures above 3000 K, there is no

choice but to use plasma. In medical settings, high temperatures and energy densities can be useful for cauterization and tissue ablation during surgery, for example.

2. Plasmas are able to produce a *very high concentration of energetic and chemically active species* (e.g., electrons, positive and negative ions, atoms and radicals, excited atoms and molecules, as well as photons that span wide spectral ranges). A high concentration of active species is crucial for important plasma applications such as plasma-assisted ignition and combustion (probably the oldest plasma application) and plasma generation of ozone for water cleaning. In medical settings, generation of the high concentration of excited and reactive species can be useful for sterilization of surfaces, air, and water streams, as well as for tissue engineering.
3. Plasma systems can be very far from thermodynamic equilibrium, providing an *extremely high concentration of the chemically active species while maintaining bulk temperatures as low as room temperature*. This feature determines exclusiveness of plasma use in microelectronics and semiconductor industries: most elements of modern computers, cell phones, television equipment, cold lighting, and other electronic devices are manufactured using cold plasma technology. This important feature also determines the wide application of cold plasma in treatment of polymers: most textiles for our clothes, photographic paper, wrapping materials and so on are today plasma treated. In medical settings, the generation of an extremely high concentration of the chemically active species, while maintaining bulk temperatures as low as room temperature, can be useful for: non-thermal blood coagulation; corrections of blood composition and properties; sterilization of skin and other living tissues; healing wounds; and treating diseases not effectively treated before.

The three specific plasma features described above permit significant intensification of traditional chemical and biochemical processes, improvements in their efficiency, and often successful stimulation of chemical and biochemical reactions that are not possible using conventional techniques.

1.3 Natural and man-made, completely and weakly ionized plasmas

Plasma comprises the majority of the mass in the known universe: the solar corona, solar wind, nebula, and the Earth's ionosphere are all plasmas. The most readily recognized form of natural plasma phenomenon in the Earth's atmosphere is lightning. The breakthrough experiments with this natural form of plasma were performed long ago by Benjamin Franklin (Figure 1.4), which explains the special interest in plasma research in the Philadelphia area where the authors of this book are based (Drexel Plasma Institute, Drexel University).

At altitudes of approximately 100 km, the atmosphere no longer remains non-conducting due to significant ionization and formation of plasma by solar radiation. As one progresses further into near-space altitudes, the Earth's magnetic field interacts with charged particles streaming from the sun. These particles are diverted and often become trapped by the Earth's magnetic field. The trapped particles are most dense near the poles, creating the beautifully rendered Aurora Borealis (Figure 1.5). Lightning and the Aurora Borealis are the most common forms of natural plasmas observed on earth.

Natural and man-made or manufactured plasmas (generated in gas discharges) occur over a wide range of pressures, electron temperatures, and electron densities (see Figure 1.6). Temperatures of manufactured plasmas range from slightly above room temperature to temperatures comparable to the interior of stars, with electron densities that span over 15 orders of magnitude. Most plasmas of practical significance, however, have electron temperatures of 1–20 eV with electron densities in the range 10^6 – 10^{18} cm^{-3} (high temperatures are conventionally expressed in electron-volts, with 1 eV \approx 11 600 K).

Not all particles need to be ionized in plasma; a common condition in plasma chemistry is for the gases to be only partially ionized. The ionization degree (ratio of density of major charged species to that of neutral gas)



Figure 1.4 Benjamin Franklin successfully performed the first experiments with the atmospheric plasma phenomenon of lightning.

in conventional plasma-chemical systems is in the range 10^{-7} – 10^{-4} . When the ionization degree is close to unity, such plasma is referred to as *completely ionized plasma*. Completely ionized plasmas are conventional for thermonuclear plasma systems (tokomaks, stellarators, plasma pinches, focuses, etc.). When ionization degree is low, the plasma is called *weakly ionized plasma*. Weakly ionized plasmas and the important chemical and biochemical processes stimulated in such plasmas is the focus of this book.



Figure 1.5 Aurora borealis.




Figure 1.6 General chart of plasma temperatures and densities.

Both natural and manufactured or man-made laboratory plasmas are quasi-neutral, which means that concentrations of positively charged particles (positive ions) and negatively charged particles (electrons and negative ions) are well balanced. Langmuir was one of the pioneers who studied gas discharges, and defined plasma to be a region not influenced by its boundaries. The transition zone between the plasma and its boundaries was termed the plasma *sheath*. The properties of the sheath differ from those of the plasma and these boundaries influence the motion of the charge particles in this sheath. They form an electrical screen for the plasma from influences of the boundary. Very important concepts group plasma physics, plasma chemistry, and plasma medicine into two major classes – those of thermal and non-thermal plasmas – which are discussed in the following section.

1.4 Plasma as a non-equilibrium multi-temperature system

Temperature in plasma is determined by the average energies of the plasma particles (neutral and charged) and their relevant degrees of freedom (translational, rotational, vibrational, and those related to electronic excitation). As multi-component systems, plasmas are therefore able to exhibit multiple temperatures. In electric discharges common for plasmas generated in the laboratory, energy from the electric field is first accumulated by the electrons through collisions; it is subsequently transferred from the electrons to the heavy particles. Electrons receive energy from the electric field during their mean free path. During the following collision with a heavy particle, they only lose a small portion of that energy (because electrons are much lighter than the heavy particles). That is why electron temperature in plasma is initially higher than that of heavy particles. Subsequently, collisions of electrons with heavy particles (Joule heating) can equilibrate their temperatures unless time or energy are not sufficient for the equilibration (such as the situation in coronas and pulsed discharges), or there is an intensive cooling mechanism preventing heating of the entire gas (as for wall-cooled low-pressure discharges).




Figure 1.7 Solar plasma.

The temperature difference between electrons and heavy neutral particles due to Joule heating in the collisional weakly ionized plasma is conventionally proportional to the square of the ratio of the electric field (E) to the pressure (p). Only in the case of small values of E/p , the temperatures of electrons and heavy particles approach each other. This is a basic requirement for the so-called local thermodynamic equilibrium (LTE) in plasma. Additionally, LTE conditions require chemical equilibrium as well as restrictions on the gradients. The LTE plasma follows major laws of the equilibrium thermodynamics and can be characterized by a single temperature at each point of space. Ionization and chemical processes in such plasmas are determined by temperature (and only indirectly by the electric fields through Joule heating). The quasi-equilibrium plasma of this kind is usually called *thermal plasma*. In nature, thermal plasmas can be represented by solar plasma (see Figure 1.7).

Numerous plasmas are sustained very far from the thermodynamic equilibrium and are characterized by multiple temperatures related to different plasma particles and different degrees of freedom. The electron temperature often significantly exceeds those of heavy particles ($T_e \gg T_0$). Ionization and chemical processes in such non-equilibrium plasmas are directly determined by electron temperature, and are therefore not very sensitive to thermal processes and temperature of the gas. The non-equilibrium plasma of this kind is usually referred to as *non-thermal plasma*. Non-thermal plasmas in nature are represented by the Aurora Borealis (see Figure 1.5) as opposed to thermal plasmas which are represented by lightning.

Although the relation between different plasma temperatures in non-thermal plasmas can be complex, it can be conventionally presented in the collisional weakly ionized plasmas as: $T_e > T_v > T_r \approx T_i \approx T_0$. Electron temperature T_e is the highest in the system, followed by the temperature of vibrational excitation of molecules T_v . The lowest temperature is usually shared in plasma by heavy neutrals (T_0 , temperature of translational degrees of freedom or simply bulk gas temperature), ions (T_i), and rotational degrees of freedom of molecules (T_r). In many non-thermal plasma systems, electron temperature is *c.* 1 eV (*c.* 10 000 K), while gas temperature is close to room temperature.

Non-thermal plasmas are usually generated either at low pressures, at lower power levels, or in a different kind of pulsed discharge systems. The engineering aspects and application realms are quite different for thermal and non-thermal plasmas. Thermal plasmas are usually more powerful (up to 30 MW and above), while non-thermal plasmas are more selective and can be used in delicate applications without degrading the surrounding environment. However, these diverse forms of ionized gases share many common characteristics.

It is interesting to note that both thermal and non-thermal plasmas usually have the highest temperature (T_e in one case and T_0 in the other) of the order of magnitude of 1 eV, which is c. 10% of the total energy required for ionization (c. 10 eV). This reflects the general axiom formulated by Zeldovich and Frank-Kamenetsky for atoms and small molecules in chemical kinetics: the temperature required for a chemical process is typically c. 10% of the total required energy, which is the Arrhenius activation energy. Plasma temperatures can be somewhat identified as the 'down payment' for the ionization process (since a similar rule i.e. that of 10% is usually applied to calculate a down payment for a mortgage).

Thermal and non-thermal plasmas have their own specific niches for biological and medical applications. High temperatures and high energy densities typical of thermal plasmas determine their applications for cauterization and tissue ablation during surgeries. Such devices are widely used today in medical practice; some of them even combine the above-mentioned features with tissue sterilization. Thermal plasma in air is also very productive in the generation of NO, which determines its application in the so-called plasma-induced NO-therapy effective in plasma treatment of wounds and different diseases. Non-thermal plasma permits the generation of an extremely high concentration of the chemically active species, while maintaining bulk temperatures as low as room temperature. It determines the specific application niche of the non-thermal plasma, which is usually: non-thermal blood coagulation and corrections of blood composition and properties; sterilization of skin and other living tissues; sterilization of medical instruments and other fragile materials and devices; processing of biopolymers; tissue engineering; and finally non-thermal plasma healing of wounds and different diseases not effectively treated before.

1.5 Gas discharges as plasma sources for biology and medicine

Plasma medicine is based on a sequence of plasma-chemical and biochemical processes involving ionized gases. A plasma source, which in most laboratory conditions is a gas discharge, therefore represents a physical and engineering basis of the plasma medicine. For simplicity, an electric discharge as a plasma source in general can be viewed as two electrodes inserted into a glass tube and connected to a power supply. The tube can be filled with various gases or evacuated. As the voltage applied across the two electrodes increases, the current suddenly increases sharply at a threshold voltage required for sufficiently intensive electron avalanches. If the pressure is low (of the order a few Torr) and the external circuit has a large resistance to prohibit a large current, a glow discharge develops. This is the low-current high-voltage discharge widely used to generate non-thermal plasmas. A similar discharge is the basis of operation for fluorescent lamps, a common plasma discharge device. The glow discharge can be considered as a major example of the low-pressure non-thermal plasma sources (see Figure 1.8). Low-pressure plasma discharges can be effective as UV sources or sources of some active species for sterilization. They can be also effective for the treatment of biopolymers and in tissue engineering. It should be mentioned that, to keep treatment areas clean from the products of erosion of electrodes, low-pressure plasma-medical technologies are often based on electrode-less low-pressure plasma sources such as low-pressure radio-frequency plasma sources.

Historically, some important developments in the area of plasma medicine probably started with the work on surface treatment and subsequent surface interactions with cells (e.g. the work carried out by the groups of Riccardo D'Agostino, Pietro Favia and Michael Wertheimer) and sterilization using low-pressure non-thermal plasma (e.g. the work carried out by the group led by Michel Moisan).

Most plasma-medical applications require operation at atmospheric pressure, therefore use of atmospheric-pressure plasma discharges. Igor Alexeff and Mounir Laroussi were some of the first researchers to employ atmospheric-pressure plasma for sterilization, while Eva Stoffels was probably one of the first to apply such discharges directly to cells. Dr. Richard Satava, who managed various projects at the Defense Advanced Research Projects Agency (DARPA), helped develop initial applications of non-thermal plasma in medicine in the United States.




Figure 1.8 *Glow discharge.*

Probably the simplest example of such discharges is the corona discharge (see Figure 1.9). A non-thermal corona discharge occurs at high pressures (including atmospheric pressure) in regions of sharply non-uniform electric fields. The field near one or both electrodes must be stronger than in the rest of the gas. This occurs near sharp points, edges or small diameter wires. These tend to be low-power plasma sources, limited by the onset of electrical breakdown of the gas. However, it is possible to circumvent this restriction through




Figure 1.9 *Corona discharge.*